

HPLC Application

ID No.: 16024

Prodigy Test Mix using Onyx C18E 100x3.0mm (0.8mL/min)

Column: Onyx[™] Monolithic C18, LC Column 100 x 3 mm, Ea

Dimensions: 100 x 3 mm ID

Order No: CH0-8158

Elution Type: Isocratic

Eluent A: Water

Eluent B: Acetonitrile


Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	35	65

Flow Rate: 0.8 mL/min

Col. Temp.: ambient

Detection: Refractive Index @ 0.000000000 (ambient)

16024



Products used in this application:



ANALYTES:

- 1 Uracil
- 2 Acetophenone
- 3 Benzene
- 4 Toluene
- 5 Naphthalene

