


25-OH Vitamin D2 and D3 Control from Serum on Kinetex 2.6µm C18 after PPT**Column:** Kinetex® 2.6 µm C18 100 Å, LC Column 50 x 4.6 mm, Ea**Dimensions:** 50 x 4.6 mm ID**Order No:** 00B-4462-E0**Elution Type:** Gradient**Eluent A:** 0.05 % Formic acid**Eluent B:** 5 mM Ammonium acetate + 0.1 % Formic acid in methanol

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	92	8
	2	0.08	92	8
	3	3.4	0	100
	4	4.8	0	100
	5	6	92	8

Flow Rate: 1 mL/min**Col. Temp.:** ambient**Detection:** Mass Spectrometer (MS) @ amu (ambient)**Detector Info:** SCIEX**Products used in this application:****ANALYTES:**

- 1 25-Hydroxy vitamin D3 (25-OH D3)
Retention Time: 0.44 min
- 2 25-Hydroxy vitamin D3-d6 (25-OH D3-d6)
Retention Time: 0.44 min
- 3 25-Hydroxy vitamin D2 (25-OH D2)
Retention Time: 0.47 min

