HPLC Application

ID No.: 19760



Kinetex[®]

Products used in this application:

Testosterone from Male Human Plasma by LC/MS/MS using Strata-X-A and Kinetex 1.7 μm C18

Kinetex® 1.7 μ m C18 100 Å, LC Column 30 x 2.1 mm, Ea

30 x 2.1 mm ID **Dimensions:** Order No: 00A-4475-AN Elution Type: Gradient

Eluent A: 0.1% Formic Acid +1 mM Amm Formate in Water **Eluent B:** 0.1% Formic Acid +1 mM Amm Formate in ACN


Liuciic Di	0.1 70 TOTTILE ACID TET THE FAITHER TOTTILE IN ACID			
Gradient	Step No.	Time (min)	Pct A	Pct B
Profile:	1	0	90	10
	2	2.5	10	90
	3	3.5	10	90
	4	3.6	90	10

Flow Rate: 0.4 mL/min Col. Temp.: ambient

Detection: Mass Spectrometer (MS) @ amu (ambient)

Detector Info: <a target="_blank"

ts/mass-spectrometers?utm_campaign=2019%20application%20search&utm_source=phenomenex&utm_medium=referral">SCIEX< 19760



ANALYTES:

Testosterone

2 Testosterone-d3 Retention Time: 2.61 min

Retention Time: 2.62 min

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com



Phenomenex products are available worldwide. www.phenomenex.com

Sample Preparation Details

for HPLC Application ID No.: 19760



Testosterone from Male Human Plasma by LC/MS/MS using Strata-X-A and Kinetex 1.7 μm C18

PRODUCT DESCRIPTION:

Strata[™]-X-A 33 µm Polymeric Strong Anion, 30 mg / 3 mL, Tubes , 50/Pk

Order No.: 8B-S123-TBJ

SOLID PHASE EXTRACTION (SPE) PRODCEDURE:

Note: The solvent volumes shown below are for a 30 mg bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

Condition:	_			
Load:	_			
Wash:	_			
Dry:				
Dry for 5 min under high vacuum	-			
Elute:	-			
Final Prep and Analysis:	_			
Following evaporation of elution solvent @ 50-55 C under gentle nitrogen stream; Add 50 uL 25% hydroxylamine solution and heat at 60-65 C for 5-10 min, then add 200 uL 5%				
Inject: 0 µL on HPLC Mass Spectrometer (MS) @ amu (ambient)				

Spiked Conc. %RSC **ANALYTES:** Log P pKa % Rec (ng/mL) (n=0)1 Testosterone 0 2 Testosterone-d3 0.5

This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com

