HPLC Application

ID No.: 19762



Testosterone from Female Human Plasma by LC/MS/MS using Strata-X-A and Kinetex 1.7 µm C18

Kinetex® 1.7 μm C18 100 Å, LC Column 30 x 2.1 mm, Ea

30 x 2.1 mm ID **Dimensions:** Order No: 00A-4475-AN Elution Type: Gradient

Eluent A: 0.1% Formic Acid +1 mM Amm Formate in Water

| Fluent B: | 0.1% Formic Acid +1 mM Amm Formate in ACN | | | | | |
|-----------|---|------------|-------|-------|--|--|
| Gradient | Step No. | Time (min) | Pct A | Pct B | | |
| Profile: | 1 | 0 | 90 | 10 | | |
| | 2 | 2.5 | 10 | 90 | | |
| | 3 | 3.5 | 10 | 90 | | |
| | 4 | 3.6 | 90 | 10 | | |

Kinetex[®]

Products used in this application:




Flow Rate: 0.4 mL/min Col. Temp.: ambient

Detection: Mass Spectrometer (MS) @ amu (ambient)

Detector Info: <a target="_blank"

-xurts/mass-spectrometers?utm_campaign=2019%20application%20search&utm_source=phenomenex&utm_medium=referral">SCIEX



ANALYTES:

2

Testosterone

Retention Time: 2.62 min Testosterone-d3 Retention Time: 2.61 min

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com



Sample Preparation Details

for HPLC Application ID No.: 19762



Testosterone from Female Human Plasma by LC/MS/MS using Strata-X-A and Kinetex 1.7 µm C18

PRODUCT DESCRIPTION:

Strata[™]-X-A 33 µm Polymeric Strong Anion, 30 mg / 3 mL, Tubes , 50/Pk

Order No.: 8B-S123-TBJ

SOLID PHASE EXTRACTION (SPE) PRODCEDURE:

Note: The solvent volumes shown below are for a 30 mg bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

| Condition: | | |
|----------------------------------|---|-----------|
| Load: | | |
| Wash: | | |
| Dry: | | |
| Dry for 5 min under high vacuum | | • |
| Elute: | | |
| Final Prep and Analysis: | | |
| 9 , | colvent @ 50-55 C under gentle nitrogen stream; olution and heat at 60-65 C for 5-10 min, then add | 200 uL 5% |
| Inject: 0 µL on HPLC Mass Spectr | ometer (MS) @ amu (ambient) | |

| ANALYTES: | Spiked Conc. | Log P | рКа | % Rec | %RSC |
|-------------------|--------------|-------|-----|-------|-------|
| | (ng/mL) | | | | (n=0) |
| 1 Testosterone | 0 | | | | |
| 2 Testosterone-d3 | 0.5 | | | | |

This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com

