

Aeris WIDEPORE 3.6u XB-C18 150 x 2.1mm ID QC Chromatogram (AL0-8931)

Column: Aeris™ 3.6 µm WIDEPORE XB-C18 200 Å, LC Column 150 x 2.1 mm, Ea

Dimensions: 150 x 2.1 mm ID

Order No: 00F-4482-AN

Elution Type: Isocratic

Eluent A: Acetonitrile

Eluent B: water

Gradient

Step No.	Time (min)	Pct A	Pct B
1	0	55	45

Profile:

Flow Rate: 0.25 mL/min


Col. Temp.: ambient

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)

Analyst Note: SecurityGuard™ Guard Cartridge System extends column lifetime.

- SecurityGuard ULTRA Cartridges, UHPLC WIDEPORE C18 for 2.1mm ID, 3/Pk Part No.: AJ0-8783


- Holder Part No.: AJ0-9000



Products used in this application:



19812



ANALYTES:

1 Uracil

Retention Time: 1.041 min

2 Acetophenone

Retention Time: 1.404 min

3 Toluene

Retention Time: 2.066 min

4 Naphthalene

Retention Time: 2.464 min

5 Acenaphthene

Retention Time: 3.543 min

