

### JHW018-hydroxypentyl isomers in Urine using Strata-X-Drug B

Kinetex® 2.6 µm C18 100 Å, LC Column 150 x 3 mm, Ea

150 x 3 mm ID **Dimensions:** Order No: 00F-4462-Y0 **Elution Type:** Gradient

Eluent A: 10mM ammonium formate

**Eluent B:** Acetonitrile

Gradient	Step No.	Time (min)	Pct A	Pct B
Profile:	1	0	55	45
	2	7	50	50
	3	7.01	5	95
	4	10	5	95

Positive


Flow Rate: 0.6 mL/min Col. Temp.: ambient

Electrospray Mass Spec (ESMS) @ 0.000000000 (ambient) **Detection:** 

Polarity: Analyst Note:

CAD: 7.00 CUR: 20.00 GS1: 50.00 GS2: 50.00 IS: 5500.00 TEM: 550.00 ihe: FΡ 10.00

App ID 20926





Products used in this application:



©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com



# **HPLC Application** ID No.: **20926**



## JHW018-hydroxypentyl isomers in Urine using Strata-X-Drug B **ANALYTES:**

- JWH-073-Butanoic acid metabolite
- JWH-018-5-Hydroxypentyl metabolite
- JWH-018-4-Hydroxypentyl metabolite



# **Sample Preparation Details**

for **HPLC** Application ID No.: **20926** 



#### JHW018-hydroxypentyl isomers in Urine using Strata-X-Drug B

#### **PRODUCT DESCRIPTION:**

Strata™-X-Drug B 33 µm Polymeric Strong Cation, 60 mg / 3 mL, Tubes , 50/Pk

Order No.: 8B-S128-UBJ

SOLID PHASE EXTRACTION (SPE) PRODCEDURE
---

<b>Note:</b> The solvent volumes shown below are for a 60 mg bed mass.
The solvent volumes will need to be adjusted for a smaller or larger bed mass.
Condition:
Load:
Sample Hydrolysis: Combine 1 mL Human Urine sample (spiked with analytes at 50 ng/mL), 2 mL of 100 mM sodium acetate buffer, pH 5.0, 25 uL -D-Glucoronidase (Patella Vulgata from Sigma, 100KU).  Vortex 10-15 secs, followed by incubation for 2 hours in a shaker at 550 C to complete hydrolysis of the glucuronides.
voitex 10-13 Secs, followed by illicubation for 2 flours in a shaker at 330 C to complete flyurolysis of the gluculorides.
Wash:
Dry:
15 mins under 10-15" of Hg
Elute:
Final Prep and Analysis:
Extraction Protocol (No need for conditioning or equilibration of the cartridge)
Inject: 10 μL on HPLC Electrospray Mass Spec (ESMS) @ 0.0000000000 (ambient)

**ANALYTES:** Spiked Conc. % Rec %RSC Log P pKa (ng/mL) (n=0)

- 1 JWH-073-Butanoic acid metabolite 0
- JWH-018-5-Hydroxypentyl metabolote
- **3** JWH-018-4-Hydroxypentyl metabol**0**te

This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com



Phenomenex products are available worldwide. www.phenomenex.com