## **HPLC Application**

ID No.: 25778



### Extraction of 5-Fluorouracil from Human Serum Using Strata-X PRO SPE on a Kinetex 2.6 µm PS C18 (ESI -Ve Mo

Kinetex® 2.6 µm PS C18 100 Å, LC Column 150 x 3 mm, Ea

**Dimensions:** 150 x 3 mm ID Order No: 00F-4780-Y0 Elution Type: Gradient

Eluent A: 0.1% Formic acid in Water

Eluent B: Methanol

Gradient Time (min) Pct A Pct B Step No. **Profile:** 0 93 7 2 12 7 93

Flow Rate: 0.2 mL/min

25 °C Col. Temp.:


**Detection:** LC/MS/MS @ 0.0000000000 m/z (mass per charge) (25 °C)

Add 100 µL Serum (doubly stripped) to tube. Add 900 µL of chilled (~00C) Acetonitrile to the serum while vortexing. Ce **Analyst Note:** 

rpm for 5 minutes and transfer supernatant. Analyte diluted with initial mobile phase.



Products used in this application:



#### **ANALYTES:**

5FU

Retention Time: 7.07 min

Retention Time: 10.82 min

5 Cl Uracil (-Ve IS) 2

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com



# Sample Preparation Details for HPLC Application ID No.: 25778



Extraction of 5-Fluorouracil from Human Serum Using Strata-X PRO SPE on a Kinetex 2.6 µm PS C18 (ESI -Ve Mo

### **PRODUCT DESCRIPTION:**

Strata X Pro, 96-Well Plate, 30mg/well

Order No.: 8E-S536-TGA

### **SOLID PHASE EXTRACTION (SPE) PRODCEDURE:**

Note: The solvent volumes shown below are for a bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

| Condition:                                       |  |
|--|--|
| Load:  |  |
| Wash:  |  |
| Dry:   |  |
| Elute:   |  |
| Final Prep and Analysis:                         |  |
| Sample Pre-treatment: • Add 100 µL Serum to tube |  |

Inject: 10 µL on HPLC LC/MS/MS @ 0.0000000000 m/z (mass per charge) (25°C)

| ANALYTES:              | Spiked Conc. | Log P | рКа | % Rec | %RSC  |
|------------------------|--------------|-------|-----|-------|-------|
|                        | (ng/mL)      |       |     |       | (n=0) |
| <b>1</b> 5FU           | 100          |       |     | 89    |       |
| 2 5 Cl Uracil (-Ve IS) | 200          |       |     |       |       |

This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com

